

FEB 03 2010

Application Serial No: 10/577,578

In response to the Notice of Non-Compliant Amendment mailed on January 27, 2010

IN THE CLAIMS

Amendments To The Claims:

This listing of claims will replace all prior versions, and listings, of claims in the application:

Listing of Claims:

1. (Currently Amended) A sample analysis method comprising: a first step of confirming that a sample is supplied to an analytical tool based on output from the analytical tool; a second step of grasping a level of the output from the analytical tool in a predetermined time period after the supply of the sample to the analytical tool is confirmed, the grasping being performed at least once including at a time point when the predetermined time period has elapsed; and a third step of performing computation necessary for the analysis of the sample;

wherein the grasping of the output from the analytical tool in the first step and the second step is performed based on output from a double integration circuit which is obtained by inputting the output from the analytical tool into the double integration circuit;

wherein the first step comprises grasping a level of the output from the double integration circuit repetitively at a first sampling time interval ~~intervals each~~ defined by a time period from ~~when start of~~ the inputting into the double integration circuit ~~is started~~ ~~till when until completion of~~ the outputting from the double integration circuit ~~is finished~~; and

wherein, ~~in grasping the second step comprises grasping~~ a level of the output from the double integration circuit ~~in the second step~~, at a second sampling time interval defined by a time period from ~~when start of~~ the inputting into the double integration circuit ~~is started till until completion of~~ ~~when the~~ outputting from the double integration circuit, the second sampling time interval being ~~is finished~~ is set longer than the first sampling time interval.

Application Serial No: 10/577,578

In response to the Notice of Non-Compliant Amendment mailed on January 27, 2010

2. (Currently Amended) The sample analysis method according to claim 1, wherein the second step comprises performing the inputting of the output from the analytical tool into the double integration circuit and the outputting from the double integration circuit repetitively at the second ~~time-intervals~~ interval, and grasping the output from the analytical tool at ~~the~~ a time point when the predetermined time period has elapsed.

3. (Currently Amended) The sample analysis method according to claim 1, wherein the first sampling time interval is selected from a range of 10 to 30 msec, whereas the second sampling time interval is selected from a range of 30 to 300 msec.

4. (Currently Amended) The sample analysis method according to claim 1, wherein the double integration circuit ~~to be used~~ includes a capacitor for storing the output from the analytical tool as an electric charge and then discharging the stored electric charge and is capable of grasping the output from the analytical tool based on ~~the~~ a discharge time of the capacitor; and

wherein a charge time of the capacitor in the first sampling time interval in the first step is shorter than a charge time of the capacitor in the second sampling time interval in the second step.

5. (Currently Amended) The sample analysis method according to claim 4, wherein the charge time of the capacitor in the first sampling time interval in the first step is selected from a range of 5 to 15 msec; and

wherein the charge time of the capacitor in the second sampling time interval in the second step is selected from a range of 15 to 150 msec.

6. (Currently Amended) The sample analysis method according to claim 1, wherein the analytical tool ~~to be used~~ includes an electrode for outputting an amount of electron transfer between the electrode and a particular component as an electric physical quantity.

Application Serial No: 10/577,578

In response to the Notice of Non-Compliant Amendment mailed on January 27, 2010

7. (Currently Amended) The sample analysis method according to claim 6, wherein the analytical tool ~~to be used~~ includes a reagent portion containing at least one reagent for promoting the electron transfer between the particular component and said electrode, and an additional electrode for applying, together with said electrode, voltage to a ~~coexistence~~ reaction system of the sample and said at least one reagent; and

wherein the electric physical quantity is outputted from the analytical tool as a current value by applying voltage to the ~~coexistence~~ reaction system using said electrode and the additional electrode.

8. (Currently Amended) The sample analysis method according to claim 1, wherein the analytical tool ~~to be used~~ is designed to use blood as the sample.

9. (Currently Amended) A sample analysis apparatus to be used with an analytical tool mounted thereto for analyzing a sample supplied to the analytical tool based on output from the analytical tool, the apparatus comprising: a double integration circuit into which the output from the analytical tool is inputted and which outputs a physical quantity related to the input; and a controller for controlling timing at which the output from the analytical tool is inputted into the double integration circuit and timing at which the physical quantity is outputted from the double integration circuit;

wherein the controller performs control so that a time interval from start of the inputting into the double integration circuit ~~til~~ until start of the outputting from the double integration circuit becomes longer ~~in a state~~ after the supply of the sample to the analytical tool is confirmed than ~~in a state~~ before the supply of the sample to the analytical tool is confirmed.

10. (Currently Amended) The sample analysis apparatus according to claim 9, wherein the double integration circuit includes a capacitor for storing the output from the analytical tool as an electric charge and then discharging the stored electric charge and is capable of grasping the output from the analytical tool based on ~~the~~ a discharge time of the capacitor; and

Application Serial No: 10/577,578

In response to the Notice of Non-Compliant Amendment mailed on January 27, 2010

wherein the controller performs control so that the charge time of the capacitor becomes longer ~~in the state~~ after the supply of the sample to the analytical tool is confirmed than ~~in the state~~ before the supply of the sample to the analytical tool is confirmed.

11. (Currently Amended) The sample analysis apparatus according to claim 10, wherein, when the analytical tool includes an electrode for outputting an electric physical quantity, the apparatus further comprises a switch for selecting a state in which the double integration circuit is directly or indirectly connected to ground or a state in which the double integration circuit is connected to the electrode; and

wherein the controller controls the switch to control the timing at which the output from the analytical tool is inputted into the double integration circuit and the timing at which the physical quantity is outputted from the double integration circuit.